



UNITED STATES PATENT AND TRADEMARK OFFICE

19
UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/519,738	12/28/2004	William Hunter Symonds	41577/308355	9233
23370	7590	06/28/2007	EXAMINER	
JOHN S. PRATT, ESQ			FOSTER, CHRISTINE E	
KILPATRICK STOCKTON, LLP			ART UNIT	PAPER NUMBER
1100 PEACHTREE STREET			1641	
ATLANTA, GA 30309				
MAIL DATE		DELIVERY MODE		
06/28/2007		PAPER		

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No.	Applicant(s)	
	10/519,738	SYMONDS, WILLIAM HUNTER	
	Examiner	Art Unit	
	Christine Foster	1641	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) Responsive to communication(s) filed on 28 February 2007.
- 2a) This action is FINAL. 2b) This action is non-final.
- 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) Claim(s) 1-8 and 10-20 is/are pending in the application.
 - 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) Claim(s) _____ is/are allowed.
- 6) Claim(s) _____ is/are rejected.
- 7) Claim(s) _____ is/are objected to.
- 8) Claim(s) 1-8 and 10-20 are subject to restriction and/or election requirement.

Application Papers

- 9) The specification is objected to by the Examiner.
- 10) The drawing(s) filed on _____ is/are: a) accepted or b) objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) All b) Some * c) None of:
 1. Certified copies of the priority documents have been received.
 2. Certified copies of the priority documents have been received in Application No. _____.
 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

1) <input type="checkbox"/> Notice of References Cited (PTO-892)	4) <input type="checkbox"/> Interview Summary (PTO-413)
2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)	Paper No(s)/Mail Date. _____
3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)	5) <input type="checkbox"/> Notice of Informal Patent Application
Paper No(s)/Mail Date _____	6) <input type="checkbox"/> Other: _____

Art Unit: 1641

DETAILED ACTION

1. Applicant's preliminary amendment, filed 12/28/04, is acknowledged and has been entered.

Election/Restrictions

2. Restriction is required under 35 U.S.C. 121 and 372.

This application contains the following inventions or groups of inventions which are not so linked as to form a single general inventive concept under PCT Rule 13.1.

In accordance with 37 CFR 1.499, applicant is required, in reply to this action, to elect a single invention to which the claims must be restricted.

Group I, claim(s) 1-8 and 10-13, drawn to a method for detecting an analyte containing a haem moiety.

Group II, claim(s) 14-18, drawn to a kit for detecting an analyte in a sample.

Group III, claim(s) 19-20, drawn to a method for detecting an analyte labeled with a haem moiety.

3. The inventions listed as Groups I-III above do not relate to a single general inventive concept under PCT Rule 13.1 because, under PCT Rule 13.2, they lack the same or corresponding special technical features for the following reasons:

According to PCT Rule 13.2, unity of invention exists only when the shared same or corresponding technical feature is a contribution over the prior art. The technical feature linking Groups I-III appears to be the features of (1) an analyte that either contains or is labeled with a heme moiety, together with (2) magnetic beads; the use of (3) luminol or other functional chemiluminescent derivative thereof; and the feature of (4) alkaline pH conditions.

Hixson et al. (WO 98/54578, Applicant's IDS of 2/28/05) teach a chemiluminescent assay for detection of the heme-containing analyte hemoglobin (the abstract and Example 2). The assay employs a binding agent bonded to the surface of a solid phase such as derivatizable paramagnetic particles (p. 2, lines 31-33; p. 6, lines 29-34), where the binding agent can be dihydroxyboryl compounds like phenylboronic acids, which selectively bind glycosylated hemoglobin, which is a heme-containing protein (p. 5, lines 14-25 and p. 8, lines 5-12). Following separation of the hemoglobin-immobilized particles, an alkaline solution (0.25M NaOH) is added (p. 8, lines 13-15). Chemiluminescent detection may be done using luminol or isoluminol for the (p. 3, lines 12-22 in particular).

Therefore, the technical feature linking the inventions of Groups I-III does not constitute a special technical feature as defined by PCT Rule 13.2, as it does not define a contribution over the prior art.

In addition, Groups I-III each have technical features that are unrelated to the other groups. Group I is a method for detecting an analyte containing a heme moiety, and includes the step of separating magnetic beads from a sample, which is not a limitation of the kit claims of Group II. Group III is a method for detecting an analyte *labeled with* a heme moiety, which is not a limitation of the method of Group II, and includes a separation step, which is not a limitation of the kit claims of Group II.

Accordingly, Groups I-III are not linked by the same or a corresponding special technical feature so as to form a single general inventive concept.

Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

The election of an invention or species may be made with or without traverse. To reserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

Should applicant traverse on the ground that the inventions or species are not patentably distinct, applicant should submit evidence or identify such evidence now of record showing the inventions or species to be obvious variants or clearly admit on the record that this is the case. In either instance, if the examiner finds one of the inventions unpatentable over the prior art, the evidence or admission may be used in a rejection under 35 U.S.C.103(a) of the other invention.

Notice of Possible Rejoinder

The examiner has required restriction between product and process claims. Where applicant elects claims directed to the product, and the product claims are subsequently found allowable, withdrawn process claims that depend from or otherwise require all the limitations of the allowable product claim will be considered for rejoinder. All claims directed a nonelected process invention must require all the limitations of an allowable product claim for that process invention to be rejoined.

In the event of rejoinder, the requirement for restriction between the product claims and the rejoined process claims will be withdrawn, and the rejoined process claims will be fully examined for

Art Unit: 1641

patentability in accordance with 37 CFR 1.104. Thus, to be allowable, the rejoined claims must meet all criteria for patentability including the requirements of 35 U.S.C. 101, 102, 103 and 112. Until all claims to the elected product are found allowable, an otherwise proper restriction requirement between product claims and process claims may be maintained. Withdrawn process claims that are not commensurate in scope with an allowable product claim will not be rejoined. See MPEP § 821.04(b). Additionally, in order to retain the right to rejoinder in accordance with the above policy, applicant is advised that the process claims should be amended during prosecution to require the limitations of the product claims. Failure to do so may result in a loss of the right to rejoinder. Further, note that the prohibition against double patenting rejections of 35 U.S.C. 121 does not apply where the restriction requirement is withdrawn by the examiner before the patent issues. See MPEP § 804.01.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christine Foster whose telephone number is (571) 272-8786. The examiner can normally be reached on M-F 8:30-5. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Long Le can be reached at (571) 272-0823. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Christine
Christine Foster, Ph.D.
Patent Examiner
Art Unit 1641

Long
LONG V. LE 06/21/07
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 1600